

2. Section 868.1150 is amended by revising paragraph (b) and by removing paragraph (c) to read as follows:

§ 868.1150 Indwelling blood carbon dioxide partial pressure (Pco₂) analyzer.

* * * * *

(b) *Classification.* Class II (special controls). The special control for this device is FDA's "Class II Special Controls Guidance Document: Indwelling Blood Gas Analyzers; Final Guidance for Industry and FDA."

3. Section 868.1170 is amended by revising paragraph (b) and by removing paragraph (c) to read as follows:

§ 868.1170 Indwelling blood hydrogen ion concentration (pH) analyzer.

* * * * *

(b) *Classification.* Class II (special controls). The special control for this device is FDA's "Class II Special Controls Guidance Document: Indwelling Blood Gas Analyzers; Final Guidance for Industry and FDA."

4. Section 868.1200 is amended by revising paragraph (b) and by removing paragraph (c) to read as follows:

§ 868.1200 Indwelling blood oxygen partial pressure (Po₂) analyzer.

* * * * *

(b) *Classification.* Class II (special controls). The special control for this device is FDA's "Class II Special Controls Guidance Document: Indwelling Blood Gas Analyzers; Final Guidance for Industry and FDA."

Dated: November 4, 2001.

Linda S. Kahan,

Deputy Director, Center for Devices and Radiological Health.

[FR Doc. 01-28561 Filed 11-14-01; 8:45 am]

BILLING CODE 4160-01-S

DEPARTMENT OF HEALTH AND HUMAN SERVICES

Food and Drug Administration

21 CFR Part 892

[Docket No. 01N-0238]

Medical Devices; Exemptions From Premarket Notification; Class II Devices

AGENCY: Food and Drug Administration, HHS.

ACTION: Final rule.

SUMMARY: The Food and Drug Administration (FDA) is publishing a final rule exempting from the premarket notification requirements the fluoroscopic compression device, a manual compression device that allows

a radiologist to press on the abdomen during a fluoroscopic procedure without exposing his or her hand to the x-ray beam. The device is classified as an accessory to the image-intensified fluoroscopic x-ray system. FDA received a petition requesting an exemption for the F-Spoon device, a type of fluoroscopic manual compression device. FDA is expanding the exemption for this type of generic device to include other fluoroscopic compression devices. FDA is publishing this order in accordance with the Food and Drug Administration Modernization Act of 1997 (FDAMA).

DATES: This rule is effective December 17, 2001.

FOR FURTHER INFORMATION CONTACT:

Heather S. Rosecrans, Center for Devices and Radiological Health (HFZ-404), Food and Drug Administration, 9200 Corporate Blvd., Rockville, MD 20850, 301-594-1190.

SUPPLEMENTARY INFORMATION:

I. Statutory Background

Under section 513 of the Federal Food, Drug, and Cosmetic Act (the act) (21 U.S.C. 360c), FDA must classify devices into one of three regulatory classes: Class I, class II, or class III. FDA classification of a device is determined by the amount of regulation necessary to provide a reasonable assurance of safety and effectiveness. Under the Medical Device Amendments of 1976 (the 1976 amendments (Public Law 94-295)), as amended by the Safe Medical Devices Act of 1990 (the SMDA (Public Law 101-629)), devices are to be classified into class I (general controls) if there is information showing that the general controls of the act are sufficient to assure safety and effectiveness; into class II (special controls), if general controls, by themselves, are insufficient to provide reasonable assurance of safety and effectiveness, but there is sufficient information to establish special controls to provide such assurance; and into class III (premarket approval), if there is insufficient information to support classifying a device into class I or class II and the device is a life-sustaining or life-supporting device or is for a use that is of substantial importance in preventing impairment of human health, or presents a potential unreasonable risk of illness or injury.

Most generic types of devices that were on the market before the date of the 1976 amendments (May 28, 1976) (generally referred to as preamendments devices) have been classified by FDA under the procedures set forth in section 513(c) and (d) of the act through the

issuance of classification regulations into one of these three regulatory classes. Devices introduced into interstate commerce for the first time on or after May 28, 1976, (generally referred to as postamendments devices) are classified through the premarket notification process under section 510(k) of the act (21 U.S.C. 360(k)). Section 510(k) of the act and the implementing regulations (21 CFR part 807) require persons who intend to market a new device to submit a premarket notification report (510(k)) containing information that allows FDA to determine whether the new device is "substantially equivalent" within the meaning of section 513(i) of the act to a legally marketed device that does not require premarket approval.

On November 21, 1997, the President signed into law FDAMA (Public Law 105-115). Section 206 of FDAMA, in part, added a new section 510(m) to the act. Section 510(m)(1) of the act requires FDA, within 60 days after enactment of FDAMA, to publish in the **Federal Register** a list of each type of class II device that does not require a report under section 510(k) of the act to provide reasonable assurance of safety and effectiveness. Section 510(m) of the act further provides that a 510(k) will no longer be required for these devices upon the date of publication of the list in the **Federal Register**. FDA published that list in the **Federal Register** of January 21, 1998 (63 FR 3142).

Section 510(m)(2) of the act provides that 1 day after date of publication of the list under section 510(m)(1) of the act, FDA may exempt a device on its own initiative or upon petition of an interested person, if FDA determines that a 510(k) is not necessary to provide reasonable assurance of the safety and effectiveness of the device. This section requires FDA to publish in the **Federal Register** a notice of intent to exempt a device, or of the petition, and to provide a 30-day comment period. Within 120 days of publication of this document, FDA must publish in the **Federal Register** its final determination regarding the exemption of the device that was the subject of the notice. If FDA fails to respond to a petition under this section within 180 days of receiving it, the petition shall be deemed granted.

II. Criteria for Exemption

There are a number of factors FDA may consider to determine whether a 510(k) is necessary to provide reasonable assurance of the safety and effectiveness of a class II device. These factors are discussed in the guidance that the agency issued on February 19, 1998, entitled "Procedures for Class II

Device Exemptions From Premarket Notification, Guidance for Industry and CDRH Staff." That guidance can be obtained through the Internet on the CDRH home page at <http://www.fda.gov/cdrh> or by facsimile through CDRH Facts-on-Demand at 1-800-899-0381 or 301-827-0111. Specify "159" when prompted for the document shelf number.

III. Petition

On March 25, 2001, FDA received a petition requesting an exemption from premarket notification for the F-Spoon, a manual compression device that allows a radiologist to press on the abdomen during a fluoroscopic procedure without exposing his or her hand to the x-ray beam. In the **Federal Register** of June 18, 2001 (66 FR 32828), FDA published a notice announcing that this petition had been received and provided an opportunity for interested persons to submit comments on the petition by July 18, 2001. FDA also announced that it intended to expand the exemption to include all fluoroscopic compression devices of this generic type, subject to limitations in 21 CFR 892.9. FDA received no comments. FDA has reviewed the petition and has determined that this device meets the criteria for exemption described previously and is, therefore, issuing this order exempting the device from the requirements of premarket notification. The fluoroscopic compression device is an accessory to the image-intensified fluoroscopic x-ray system classified under 21 CFR 892.1650.

IV. Environmental Impact

The agency has determined under 21 CFR 25.30(h) that this action is of a type that does not individually or cumulatively have a significant effect on the human environment. Therefore, neither an environmental assessment nor an environmental impact statement is required.

V. Analysis of Impacts

FDA has examined the impacts of the final rule under Executive Order 12866 and the Regulatory Flexibility Act (5 U.S.C. 601-612) (as amended by subtitle D of the Small Business Regulatory Fairness Act of 1996 (Public Law 104-121)), and the Unfunded Mandates Reform Act of 1995 (Public Law 104-4). Executive Order 12866 directs agencies to assess all costs and benefits of available regulatory alternatives and, when regulation is necessary, to select regulatory approaches that maximize net benefits (including potential economic, environmental, public health

and safety, and other advantages; distributive impacts; and equity). The agency believes that this final rule is consistent with the regulatory philosophy and principles identified in the Executive order. In addition, the final rule is not a significant regulatory action as defined by the Executive order and so is not subject to review under the Executive order.

The Regulatory Flexibility Act requires agencies to analyze regulatory options that would minimize any significant impact of a rule on small entities. Because this rule will relieve a burden and simplify the marketing of these devices, the agency certifies that the final rule will not have a significant economic impact on a substantial number of small entities. Therefore, under the Regulatory Flexibility Act, no further analysis is required.

VI. Paperwork Reduction Act of 1995

FDA concludes that this final rule contains no collections of information. Therefore, clearance by the Office of Management and Budget under the Paperwork Reduction Act of 1995 is not required.

VII. Federalism

FDA has analyzed this final rule in accordance with the principles set forth in Executive Order 13132. FDA has determined that the rule does not contain policies that have substantial direct effects on the States, on the relationship between the National Government and the States, or on the distribution of power and responsibilities among the various levels of government. Accordingly, the agency has concluded that the rule does not contain policies that have federalism implications as defined in the order and, consequently, a federalism summary impact statement is not required.

List of Subjects in 21 CFR Part 892

Medical devices, Radiation protection, X-rays.

Therefore, under the Federal Food, Drug, and Cosmetic Act and under authority delegated to the Commissioner of Food and Drugs, 21 CFR part 892 is amended as follows:

PART 892—RADIOLOGY DEVICES

1. The authority citation for 21 CFR part 892 continues to read as follows:

Authority: 21 U.S.C. 351, 360, 360c, 360e, 360j, 371.

2. Section 892.1650 is amended by revising paragraph (b) to read as follows:

§ 892.1650 Image-intensified fluoroscopic x-ray system.

* * * * *

(b) *Classification.* Class II. When intended as an accessory to the device described in paragraph (a) of this section, the fluoroscopic compression device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter subject to § 892.9.

Dated: October 25, 2001.

Linda S. Kahan,

Deputy Director, Center for Devices and Radiological Health.

[FR Doc. 01-28563 Filed 11-14-01; 8:45 am]

BILLING CODE 4160-01-S

PENSION BENEFIT GUARANTY CORPORATION

29 CFR Parts 4022 and 4044

Benefits Payable in Terminated Single-Employer Plans; Allocation of Assets in Single-Employer Plans; Interest Assumptions for Valuing and Paying Benefits

AGENCY: Pension Benefit Guaranty Corporation.

ACTION: Final rule.

SUMMARY: The Pension Benefit Guaranty Corporation's regulations on Benefits Payable in Terminated Single-Employer Plans and Allocation of Assets in Single-Employer Plans prescribe interest assumptions for valuing and paying benefits under terminating single-employer plans. This final rule amends the regulations to adopt interest assumptions for plans with valuation dates in December 2001. Interest assumptions are also published on the PBGC's Web site (<http://www.pbgc.gov>).

EFFECTIVE DATE: December 1, 2001.

FOR FURTHER INFORMATION CONTACT: Harold J. Ashner, Assistant General Counsel, Office of the General Counsel, Pension Benefit Guaranty Corporation, 1200 K Street, NW., Washington, DC 20005, 202-326-4024. (TTY/TDD users may call the Federal relay service toll-free at 1-800-877-8339 and ask to be connected to 202-326-4024.)

SUPPLEMENTARY INFORMATION: The PBGC's regulations prescribe actuarial assumptions—including interest assumptions—for valuing and paying plan benefits of terminating single-employer plans covered by title IV of the Employee Retirement Income Security Act of 1974. The interest assumptions are intended to reflect current conditions in the financial and annuity markets.