

**§ 95.1225**

**47 CFR Ch. I (10–1–14 Edition)**

adjusted using generally accepted engineering practices and standards that are mutually agreeable to both coordinators to take into account the local conditions and operating characteristics of the applicable AMT and MBAN facilities, and shall specify when the device shall limit its transmissions to segments of the 2360–2390 MHz band or shall cease operation in the band. This ITU document is incorporated by reference in accordance with 5 U.S.C. 552(a) and 1 CFR part 51 and approved by the Director of Federal Register. Copies of the recommendation may be obtained from ITU, Place des Nations, 1211 Geneva 20, Switzerland, or online at <http://www.itu.int/en/publications/Pages/default.aspx>. You may inspect a copy at the Federal Communications Commission, 445 12th Street, SW., Washington, DC 20554, or at the National Archives and Records Administration (NARA). For information on the availability of this material at NARA, call 202-741-6030, or go to: [http://www.archives.gov/federal\\_register/code\\_of\\_federal\\_regulations/ibr\\_locations.html](http://www.archives.gov/federal_register/code_of_federal_regulations/ibr_locations.html). “Generally accepted engineering practices and standards” include, but are not limited to, engineering analyses and measurement data as well as limiting MBAN operations in the band by time or frequency.

(3) If an AMT operator plans to operate a receive site not previously analyzed by the MBAN coordinator to determine line of sight to an MBAN facility, the AMT operator shall consider using locations that are beyond the line of sight of a registered health care facility. If the AMT operator determines that non-line of sight locations are not practical for its purposes, the AMT coordinator shall notify the MBAN coordinator upon no less than 7 days’ notice that the registered health care facility must cease MBAN operations in the 2360–2390 MHz band unless the parties can achieve a mutually satisfactory coordination agreement under paragraph (c)(2) of this section.

[77 FR 55734, Sept. 11, 2012]

**§ 95.1225 Frequency coordinator.**

(a) The Commission will designate a frequency coordinator(s) to manage the

operation of medical body area networks in the 2360 MHz -2390 MHz band.

(b) The frequency coordinator shall perform the following functions:

(1) Register health care facilities that operate an MBAN in the 2360–2390 MHz band, maintain a database of these MBAN transmitter locations and operational parameters, and provide the Commission with information contained in the database upon request;

(2) Determine if an MBAN is within line of sight of an AMT receive facility in the 2360–2390 MHz band and coordinate MBAN operations with the designated AMT coordinator as specified in § 87.305 of this chapter;

(3) Notify a registered health care facility when an MBAN has to change frequency within the 2360–2390 MHz band or to cease operating in the band consistent with a coordination agreement between the MBAN and the AMT coordinators;

(4) Develop procedures to ensure that registered health care facilities operate an MBAN consistent with the coordination requirements under § 95.1223; and

(5) Identify the MBAN that is the source of interference in response to a complaint from the AMT coordinator and notify the health care facility of alternative frequencies available for MBAN use or to cease operation consistent with the rules.

[77 FR 55735, Sept. 11, 2012]

**Subpart J—Multi-Use Radio Service (MURS)**

SOURCE: 65 FR 60878, Oct. 13, 2000, unless otherwise noted.

**GENERAL PROVISIONS**

**§ 95.1301 Eligibility.**

An entity is authorized by rule to operate a MURS transmitter if it is not a foreign government or a representative of a foreign government and if it uses the transmitter in accordance with § 95.1309 and otherwise operates in accordance with the rules contained in this subpart. No license will be issued.

**§ 95.1303 Authorized locations.**

(a) MURS operation is authorized: