

**§ 95.628**

radiated field strength at 3 meters for 25 microwatts, 250 nanowatts, and 100 nanowatts EIRP is 18.2, 1.8, or 1.2 mV/meter, respectively, when measured on an open area test site; or 9.1, 0.9, or 0.6 mV/meter, respectively, when measured on a test site equivalent to free space such as a fully anechoic test chamber. Compliance with the maximum transmitter power requirements set forth in §95.639(f) shall be based on measurements using a peak detector function and measured over an interval of time when transmission is continuous and at its maximum power level. In lieu of using a peak detector function, measurement procedures that have been found to be acceptable to the Commission in accordance with §2.947 of this chapter may be used to demonstrate compliance.

(2) Frequency stability testing shall be performed over the temperature range set forth in (e) of this section.

(3) Radiated emissions and EIRP measurements may be determined by measuring the radiated field from the equipment under test at 3 meters and calculating the EIRP. The equivalent radiated field strength at 3 meters for 25 microwatts, 250 nanowatts, and 100 nanowatts EIRP is 18.2, 1.8, or 1.2 mV/meter, respectively, when measured on an open area test site; or 9.1, 0.9, or 0.6 mV/meter, respectively, when measured on a test site equivalent to free space such as a fully anechoic test chamber. Compliance with the maximum transmitter power requirements set forth in §95.639(f) shall be based on measurements using a peak detector function and measured over an interval of time when transmission is continuous and at its maximum power level. In lieu of using a peak detector function, measurement procedures that have been found to be acceptable to the Commission in accordance with §2.947 of this chapter may be used to demonstrate compliance.

(i) For a transmitter intended to be implanted in a human body, radiated emissions and EIRP measurements for transmissions by stations authorized under this section may be made in accordance with a Commission-approved human body simulator and test technique. A formula for a suitable tissue

**47 CFR Ch. I (10–1–14 Edition)**

substitute material is defined in OET Bulletin 65 Supplement C (01–01).

[74 FR 22705, May 14, 2009, as amended at 75 FR 52477, Aug. 26, 2010. Redesignated and amended at 77 FR 4268, Jan. 27, 2012]

**§ 95.628 MedRadio transmitters in the 413–419 MHz, 426–432 MHz, 438–444 MHz, and 451–457 MHz and 2360–2400 MHz bands.**

The following provisions apply to MedRadio transmitters operating in the 413–419 MHz, 426–432 MHz, 438–444 MHz, and 451–457 MHz bands as part of a Medical Micropower Network (MMN) and in the 2360–2400 MHz band as part of a Medical Body Area Network (MBAN).

(a) *Operating frequencies.* A MedRadio station authorized under this part must have out-of-band emissions that are attenuated in accordance with §95.635.

(1) Only MedRadio stations that are part of an MMN may operate in the 413–419 MHz, 426–432 MHz, 438–444 MHz, and 451–457 MHz frequency bands. Each MedRadio station that is part of an MMN must be capable of operating in each of the following frequency bands: 413–419 MHz, 426–432 MHz, 438–444 MHz, and 451–457 MHz. All MedRadio stations that are part of a single MMN must operate in the same frequency band.

(2) Only MedRadio stations that are part of an MBAN may operate in the 2360–2400 MHz frequency band.

(b) *Requirements for a Medical Micropower Network.* (1) *Frequency monitoring.* MedRadio programmer/control transmitters must incorporate a mechanism for monitoring the authorized bandwidth of the frequency band that the MedRadio transmitters intend to occupy. The monitoring system antenna shall be the antenna used by the programmer/control transmitter for a communications session.

(i) The MedRadio programmer/control transmitter shall be capable of monitoring any occupied frequency band at least once every second and monitoring alternate frequency bands within two seconds prior to executing a change to an alternate frequency band.

(ii) The MedRadio programmer/control transmitter shall move to another frequency band within one second of detecting a persistent (*i.e.*, lasting

more than 50 milliseconds in duration) signal level greater than  $-60$  dBm as received by a 0 dBi gain antenna in any 12.5 kHz bandwidth within the authorized bandwidth.

(iii) The MedRadio programmer/control transmitter shall be capable of monitoring the authorized bandwidth of the occupied frequency band to determine whether either direction of the communications link is becoming degraded to the extent that communications is likely to be lost for more than 45 milliseconds. Upon making such a determination the MedRadio programmer/control transmitter shall move to another frequency band.

(2) *MedRadio transmitters.* MedRadio transmitters shall incorporate a programmable means to implement a system shutdown process in the event of communication failure, on command from the MedRadio programmer/control transmitter, or when no frequency band is available. The shutdown process shall commence within 45 milliseconds after loss of the communication link or receipt of the shutdown command from the MedRadio programmer/control transmitter.

(3) *MedRadio programmer/control transmitters.* MedRadio programmer/control transmitters shall have the ability to operate in the presence of other primary and secondary users in the 413–419 MHz, 426–432 MHz, 438–444 MHz, and 451–457 MHz bands.

(4) *Authorized bandwidth.* The 20 dB authorized bandwidth of the emission from a MedRadio station operating in the 413–419 MHz, 426–432 MHz, 438–444 MHz, and 451–457 MHz bands shall not exceed 6 MHz.

(c) *Requirements for Medical Body Area Networks.* A MedRadio programmer/control transmitter shall not commence operating and shall automatically cease operating in the 2360–2390 MHz band if it does not receive, in accordance with the protocols specified by the manufacturer, a control message permitting such operation. Additionally, a MedRadio programmer/control transmitter operating in the 2360–2390 MHz band shall comply with a control message that notifies the device to limit its transmissions to segments of the 2360–2390 MHz band or to cease operation in the band.

(d) *Frequency stability.* Each transmitter in the MedRadio service must maintain a frequency stability of  $\pm 100$  ppm of the operating frequency over the range:

(1) 25 °C to 45 °C in the case of medical implant transmitters; and

(2) 0 °C to 55 °C in the case of MedRadio programmer/control transmitters and Medical body-worn transmitters.

(e) *Shared access.* The provisions of this section shall not be used to extend the range of spectrum occupied over space or time for the purpose of denying fair access to spectrum for other MedRadio systems.

(f) *Measurement procedures.* (1) MedRadio transmitters shall be tested for frequency stability, radiated emissions and EIRP limit compliance in accordance with paragraphs (f)(2) and (3) of this section.

(2) Frequency stability testing shall be performed over the temperature range set forth in (d) of this section.

(3) Radiated emissions and EIRP limit measurements may be determined by measuring the radiated field from the equipment under test at 3 meters and calculating the EIRP. The equivalent radiated field strength at 3 meters for 1 milliwatt, 25 microwatts, 250 nanowatts, and 100 nanowatts EIRP is 115.1, 18.2, 1.8, or 1.2 mV/meter, respectively, when measured on an open area test site; or 57.55, 9.1, 0.9, or 0.6 mV/meter, respectively, when measured on a test site equivalent to free space such as a fully anechoic test chamber. Compliance with the maximum transmitter power requirements set forth in § 95.639(f) shall be based on measurements using a peak detector function and measured over an interval of time when transmission is continuous and at its maximum power level. In lieu of using a peak detector function, measurement procedures that have been found to be acceptable to the Commission in accordance with § 2.947 of this chapter may be used to demonstrate compliance. For a transmitter intended to be implanted in a human body, radiated emissions and EIRP measurements for transmissions by stations authorized under this section may be made in accordance with a Commission-approved human body

§ 95.629

simulator and test technique. A formula for a suitable tissue substitute material is defined in OET Bulletin 65 Supplement C (01-01).

[77 FR 55732, Sept. 11, 2012]

§ 95.629 LPRS transmitter frequencies.

(a) LPRS transmitters may operate on any frequency listed in paragraphs (b), (c), and (d) of this section. Channels 19, 20, 50, and 151-160 are available exclusively for law enforcement tracking purposes. AMTS transmissions are limited to the 216.750-217.000 MHz band for low power point-to-point network control communications by AMTS coast stations. Other AMTS transmissions in the 216-217 MHz band are prohibited.

(b) *Standard band channels.* (1) The following table indicates standard band frequencies. The channel bandwidth is 25 kHz.

Channel No.	Center frequency (MHz)
1	216.0125
2	216.0375
3	216.0625
4	216.0875
5	216.1125
6	216.1375
7	216.1625
8	216.1875
9	216.2125
10	216.2375
11	216.2625
12	216.2875
13	216.3125
14	216.3375
15	216.3625
16	216.3875
17	216.4125
18	216.4375
19	216.4625
20	216.4875
21	216.5125
22	216.5375
23	216.5625
24	216.5875
25	216.6125
26	216.6375
27	216.6625
28	216.6875
29	216.7125
30	216.7375
31	216.7625
32	216.7875
33	216.8125
34	216.8375
35	216.8625
36	216.8875
37	216.9125
38	216.9375
39	216.9625
40	216.9875

47 CFR Ch. I (10-1-14 Edition)

(2) LPRS transmitters operating on standard band channels must be maintained within a frequency stability of 50 parts per million.

(c) *Extra band channels.* (1) The following table indicates extra band frequencies. The channel bandwidth is 50 kHz.

Channel No.	Center frequency (MHz)
41	216.025
42	216.075
43	216.125
44	216.175
45	216.225
46	216.275
47	216.325
48	216.375
49	216.425
50	216.475
51	216.525
52	216.575
53	216.625
54	216.675
55	216.725
56	216.775
57	216.825
58	216.875
59	216.925
60	216.975

(2) LPRS transmitters operating on extra band channels must be maintained within a frequency stability of 50 parts per million.

(d) *Narrowband channels.* (1) The following table indicates narrowband frequencies. The channel bandwidth is 5 kHz and the authorized bandwidth is 4 kHz.

Channel No.	Center frequency (MHz)
61	216.0025
62	216.0075
63	216.0125
64	216.0175
65	216.0225
66	216.0275
67	216.0325
68	216.0375
69	216.0425
70	216.0475
71	216.0525
72	216.0575
73	216.0625
74	216.0675
75	216.0725
76	216.0775
77	216.0825
78	216.0875
79	216.0925
80	216.0975
81	216.1025
82	216.1075
83	216.1125
84	216.1175