

and beds for use beyond a 24 hour period in rendering medical treatment.

[79 FR 60099, Oct. 6, 2014]

§ 95.1205 Station identification.

A station is not required to transmit a station identification announcement.

§ 95.1207 Station inspection.

Any non-implanted MedRadio transmitter must be made available for inspection upon request by an authorized FCC representative. Persons operating implanted or body-worn MedRadio transmitters shall cooperate reasonably with duly authorized FCC representatives in the resolution of interference.

§ 95.1209 Permissible communications.

(a) Except for the purposes of testing and for demonstrations to health care professionals, MedRadio programmer/control transmitters may transmit only non-voice data containing operational, diagnostic and therapeutic information associated with a medical implant device or medical body-worn device that has been implanted or placed on the person by or under the direction of a duly authorized health care professional.

(b) Except as provided in § 95.627(b) no MedRadio implant or body-worn transmitter shall transmit except in response to a transmission from a MedRadio programmer/control transmitter or in response to a non-radio frequency actuation signal generated by a device external to the body with respect to which the MedRadio implant or body-worn transmitter is used.

(c) MedRadio programmer/control transmitters may be interconnected with other telecommunications systems including the public switched telephone network.

(d) For the purpose of facilitating MedRadio system operation during a MedRadio communications session, as defined in § 95.627, MedRadio transmitters in the 401–406 MHz band may transmit in accordance with the provisions of § 95.627(a) for no more than 5 seconds without the communications of data; MedRadio transmitters may transmit in accordance with the provisions of § 95.627(b)(2) and (b)(3) for no more than

3.6 seconds in total within a one hour time period; and MedRadio transmitters may transmit in accordance with the provisions of § 95.627(b)(4) for no more than 360 milliseconds in total within a one hour time period.

(e) MedRadio programmer/control transmitters may not be used to relay information in the 401–406 MHz band to a receiver that is not included with a medical implant or medical body-worn device. Wireless retransmission of information intended to be transmitted by a MedRadio programmer/control transmitter or information received from a medical implant or medical body-worn transmitter shall be performed using other radio services that operate in spectrum outside of the 401–406 MHz band.

(f) MedRadio programmer/control transmitters and medical implant transmitters may not be used to relay information in the 413–419 MHz, 426–432 MHz, 438–444 MHz, and 451–457 MHz bands to a receiver that is not part of the same Medical Micropower Network. Wireless retransmission of information to a receiver that is not part of the same Medical Micropower Network must be performed using other radio services that operate in spectrum outside of the 413–419 MHz, 426–432 MHz, 438–444 MHz, and 451–457 MHz bands. Notwithstanding the above restrictions, a MedRadio programmer/control transmitter of an MMN may communicate with the MedRadio programmer/control transmitter of another MMN to coordinate transmissions so as to avoid interference between the two MMNs.

(g) Medical body-worn transmitters may relay only information in the 2360–2400 MHz band to a MedRadio programmer/control transmitter or another medical body-worn transmitter device that is part of the same Medical Body Area Network (MBAN). A MedRadio programmer/control transmitter may not be used to relay information in the 2360–2400 MHz band to other MedRadio programmer/controller transmitters. Wireless retransmission of all other information from an MBAN transmitter to a receiver that is not part of the same MBAN shall be performed using other radio services that operate in spectrum outside of the 2360–2400 MHz band. Notwithstanding