

§ 95.2595

47 CFR Ch. I (10–1–20 Edition)

This device may not interfere with stations operating in the 400.150–406.000 MHz band in the Meteorological Aids, Meteorological Satellite, and Earth Exploration Satellite Services and must accept any interference received, including interference that may cause undesired operation.

(b) MedRadio programmer/control transmitters operating in the 413–419 MHz, 426–432 MHz, 438–444 MHz, and 451–457 MHz bands shall be labeled as provided in part 2 of this chapter and shall bear the following statement in a conspicuous location on the device:

This device may not interfere with stations authorized to operate on a primary basis in the 413–419 MHz, 426–432 MHz, 438–444 MHz, and 451–457 MHz bands, and must accept any interference received, including interference that may cause undesired operation.

(c) MedRadio programmer/control transmitters operating in the 2360–2400 MHz band shall be labeled as provided in part 2 of this chapter and shall bear the following statement in a conspicuous location on the device:

This device may not interfere with stations authorized to operate on a primary basis in the 2360–2400 MHz band, and must accept any interference received, including interference that may cause undesired operation.

(d) If it is not feasible to place the statement specified by paragraph (a), (b), or (c) of this section on the device, it may be placed in the instruction manual for the transmitter instead.

(e) If a MedRadio programmer/control transmitter is constructed in two or more sections connected by wire and marketed together, the statement specified in this section is required to be affixed only to the main control unit.

(f) MedRadio transmitters shall be identified with a serial number on each device, except as noted in paragraphs (f)(1) and (2) of this section.

(1) For MedRadio transmitters that operate in the 2360–2400 MHz band, only the programmer/control transmitter shall be identified with a serial number.

(2) The FCC ID number associated with a medical implant transmitter and the information required by § 2.925 of this chapter may be placed in the instruction manual for the transmitter

and on the shipping container for the transmitter, in lieu of being placed directly on the transmitter.

§ 95.2595 MedRadio disclosures.

Manufacturers of MedRadio transmitters must include with each transmitting device the statement set forth in this section that applies to the frequency bands in use.

(a) For MedRadio transmitters operating in the 401–406 MHz band, the following statement applies:

This transmitter is authorized by rule under the Medical Device Radiocommunication Service (in part 95 of the FCC Rules) and must not cause harmful interference to stations operating in the 400.150–406.000 MHz band in the Meteorological Aids (*i.e.*, transmitters and receivers used to communicate weather data), the Meteorological Satellite, or the Earth Exploration Satellite Services and must accept interference that may be caused by such stations, including interference that may cause undesired operation. This transmitter shall be used only in accordance with the FCC Rules governing the Medical Device Radiocommunication Service. Analog and digital voice communications are prohibited. Although this transmitter has been approved by the Federal Communications Commission, there is no guarantee that it will not receive interference or that any particular transmission from this transmitter will be free from interference.

(b) For MedRadio transmitters operating in the 413–419 MHz, 426–432 MHz, 438–444 MHz and 451–457 MHz bands, the following statement applies:

This transmitter is authorized by rule under the MedRadio Service (47 CFR part 95). This transmitter must not cause harmful interference to stations authorized to operate on a primary basis in the 413–419 MHz, 426–432 MHz, 438–444 MHz, and 451–457 MHz bands, and must accept interference that may be caused by such stations, including interference that may cause undesired operation. This transmitter shall be used only in accordance with the FCC Rules governing the MedRadio Service. Analog and digital voice communications are prohibited. Although this transmitter has been approved by the Federal Communications Commission, there is no guarantee that it will not receive interference or that any particular transmission from this transmitter will be free from interference.

(c) For MedRadio transmitters operating in the 2360–2400 MHz band, the following statement applies: