

Seven copies of the request should be submitted to the Assistant Secretary for Import Administration, International Trade Administration, Room B-099, U.S. Department of Commerce, 14th Street & Constitution Avenue, NW, Washington, DC 20230. The Department also asks parties to serve a copy of their requests to the Office of Antidumping and Countervailing Duty Enforcement, Attention: Sheila Forbes, in Room 3064 of the main Commerce building. Further, in accordance with section 353.31(g) of the regulations, a copy of each request must be served on every party on the Department's service list.

The Department will publish in the Federal Register a notice of "Initiation of Antidumping Duty Administrative Review," for requests received by October 31, 1996. If the Department does not receive by October 31, 1996 a request for review of entries covered by the order or finding listed in this notice and for the period identified above, the Department will instruct the Customs Service to assess antidumping duties on those entries at a rate equal to the cash deposit of, or bond for, estimated antidumping duties required on those entries at the time of entry, or withdrawal from warehouse, for consumption, and to continue to collect the cash deposit previously ordered.

This notice is not required by the statute, but is published as a service to the international trading community.

Dated: October 23, 1996.

Roland L. MacDonald,

Acting Deputy Assistant Secretary for Enforcement Group III.

[FR Doc. 96-27770 Filed 10-29-96; 8:45 am]

BILLING CODE 3510-DS-P

International Trade Administration

Cornell University; Notice of Decision on Application for Duty-Free Entry of Scientific Instrument

This decision is made pursuant to Section 6(c) of the Educational, Scientific, and Cultural Materials Importation Act of 1966 (Pub. L. 89-651, 80 Stat. 897; 15 CFR part 301). Related records can be viewed between 8:30 A.M. and 5:00 P.M. in Room 4211, U.S. Department of Commerce, 14th and Constitution Avenue, N.W., Washington, D.C.

Docket Number: 96-087. *Applicant:* Cornell University, Ithaca, NY 14853. *Instrument:* Scanning Tunneling Microscope, Model JSTM-4500. *Manufacturer:* JEOL Ltd., Japan. *Intended Use:* See notice at 61 FR 46783, September 5, 1996.

Comments: None received. *Decision:* Approved. No instrument of equivalent scientific value to the foreign instrument, for such purposes as it is intended to be used, is being manufactured in the United States. *Reasons:* The foreign instrument provides: (1) an ultra-high vacuum STM chamber operable to 2×10^{-8} Pa or less and (2) resolution of 0.14 nm (horizontal) with drift ≤ 0.05 nm/s at a sample temperature of 30K. A National Science Foundation engineering research center advises that (1) these capabilities are pertinent to the applicant's intended purpose and (2) it knows of no domestic instrument or apparatus of equivalent scientific value to the foreign instrument for the applicant's intended use.

We know of no other instrument or apparatus of equivalent scientific value to the foreign instrument which is being manufactured in the United States.

Frank W. Creel,

Director, Statutory Import Programs Staff.

[FR Doc. 96-27774 Filed 10-29-96; 8:45 am]

BILLING CODE 3510-DS-P

International Trade Administration

Mayo Foundation; Notice of Decision on Application for Duty-Free Entry of Scientific Instrument

This decision is made pursuant to Section 6(c) of the Educational, Scientific, and Cultural Materials Importation Act of 1966 (Pub. L. 89-651, 80 Stat. 897; 15 CFR part 301). Related records can be viewed between 8:30 A.M. and 5:00 P.M. in Room 4211, U.S. Department of Commerce, 14th and Constitution Avenue, N.W., Washington, D.C.

Docket Number: 96-084. *Applicant:* Mayo Foundation, Rochester, MN 55905. *Instrument:* IR Mass Spectrometer with Gas Sampling Inlet, Model TracerMAT. *Manufacturer:* Finnigan MAT, Germany. *Intended Use:* See notice at 61 FR 46782, September 5, 1996.

Comments: None received. *Decision:* Approved. No instrument of equivalent scientific value to the foreign instrument, for such purposes as it is intended to be used, is being manufactured in the United States.

Reasons: The foreign instrument provides: (1) a magnetic sector analyzer with three Faraday collectors tuned to isotopically labelled CO_2 , (2) an autosampler gas chromatograph designed specifically to separate CO_2 from other gases in breath samples and (3) a precision of 0.3 per mil. Two domestic manufacturers of similar

equipment advise that (1) these capabilities are pertinent to the applicant's intended purpose and (2) they know of no domestic instrument or apparatus of equivalent scientific value to the foreign instrument for the applicant's intended use.

We know of no other instrument or apparatus of equivalent scientific value to the foreign instrument which is being manufactured in the United States.

Frank W. Creel,

Director, Statutory Import Programs Staff.

[FR Doc. 96-27861 Filed 10-29-96; 8:45 am]

BILLING CODE 3510-DS-P

National Institutes of Health, et al.; Notice of Consolidated Decision on Applications for Duty-Free Entry of Electron Microscopes

This is a decision consolidated pursuant to Section 6(c) of the Educational, Scientific, and Cultural Materials Importation Act of 1966 (Pub. L. 89-651, 80 Stat. 897; 15 CFR part 301). Related records can be viewed between 8:30 A.M. and 5:00 P.M. in Room 4211, U.S. Department of Commerce, 14th and Constitution Avenue, N.W., Washington, D.C.

Docket Number: 96-085. *Applicant:* National Institutes of Health, Bethesda, MD 20892. *Instrument:* Electron Microscope, Model CM 120. *Manufacturer:* Philips, The Netherlands. *Intended Use:* See notice at 61 FR 46782, September 5, 1996. *Order Date:* March 5, 1996.

Docket Number: 96-088. *Applicant:* The University of Texas at Austin, Austin, TX 78712. *Instrument:* Electron Microscope, Model JEM-2010. *Manufacturer:* JEOL Ltd., Japan. *Intended Use:* See notice at 61 FR 46783, September 5, 1996. *Order Date:* September 30, 1993.

Docket Number: 96-093. *Applicant:* The Ohio State University, Columbus, OH 43210. *Instrument:* Electron Microscope, Model CM300. *Manufacturer:* Philips, The Netherlands. *Intended Use:* See notice at 61 FR 49113, September 18, 1996. *Order Date:* December 5, 1995.

Comments: None received. *Decision:* Approved. No instrument of equivalent scientific value to the foreign instrument, for such purposes as these instruments are intended to be used, was being manufactured in the United States at the time the instruments were ordered. *Reasons:* Each foreign instrument is a conventional transmission electron microscope (CTEM) and is intended for research or scientific educational uses requiring a CTEM. We know of no CTEM, or any

other instrument suited to these purposes, which was being manufactured in the United States either at the time of order of each instrument or at the time of receipt of application by the U.S. Customs Service.

Frank W. Creel,

Director, Statutory Import Programs Staff.

[FR Doc. 96-27773 Filed 10-29-96; 8:45 am]

BILLING CODE 3510-DS-P

Applications for Duty-Free Entry of Scientific Instruments

Pursuant to Section 6(c) of the Educational, Scientific and Cultural Materials Importation Act of 1966 (Pub. L. 89-651; 80 Stat. 897; 15 CFR part 301), we invite comments on the question of whether instruments of equivalent scientific value, for the purposes for which the instruments shown below are intended to be used, are being manufactured in the United States.

Comments must comply with 15 CFR 301.5(a)(3) and (4) of the regulations and be filed within 20 days with the Statutory Import Programs Staff, U.S. Department of Commerce, Washington, D.C. 20230. Applications may be examined between 8:30 A.M. and 5:00 P.M. in Room 4211, U.S. Department of Commerce, 14th Street and Constitution Avenue, N.W., Washington, D.C.

Docket Number: 95-080R. Applicant: Santa Rosa Outpatient Rehabilitation Hospital, 2829 Babcock Road, San Antonio, TX 78229. Instrument: 3-Dimensional Motion Analyzer System, Model VICON 370. Manufacturer: Oxford Metrics, Ltd., United Kingdom. Intended Use: Original notice of this resubmitted application was published in the Federal Register of September 19, 1995.

Docket Number: 96-102. Applicant: Yale University, Magnetic Resonance Center, 333 Cedar Street, P. O. Box 208043, New Haven, CT 06520. Instrument: SIMS IVS Console. Manufacturer: Surrey Medical Imaging Systems Ltd., United Kingdom. Intended Use: The instrument will be used to develop and apply magnetic resonance methods for imaging blood flow, tissue perfusion, intra and extracellular swelling, alterations in cellular membranes, tissue fuel sources, metabolic fuel consumption, enzymatic regulation of metabolism by using an existing 4.7 Tesla magnetic resonance spectrometer. Application accepted by Commissioner of Customs: September 27, 1996.

Docket Number: 96-103. Applicant: Stevens Institute of Technology, Castle

Point on Hudson, Hoboken, NJ 07030. Instrument: Stopped-Flow/Scanning Spectrometer, Model SX.18MV. Manufacturer: Applied Photophysics Ltd., United Kingdom. Intended Use: The instrument will be used for studies of the kinetics of human alcohol dehydrogenase isoenzymes from the liver and stomach and for studies of the kinetics of a human liver cytochrome P450 isoenzyme that metabolizes ethanol. Application accepted by Commissioner of Customs: October 1, 1996.

Docket Number: 96-104. Applicant: University of Georgia, D W Brooks Drive, Warnell School of Forest Resources, Building #4, Room 102, Athens, GA 30602. Instrument: Environmental Process Control Laboratory. Manufacturer: Minworth Systems Ltd., United Kingdom. Intended Use: The instrument will be used to monitor the transport and biochemical transformation of carbon-, nitrogen- and phosphorus-bearing materials in water and the behavior of the microbiological organisms responsible for these biochemical transformations. The goal of the research is to support the development and evaluation of computer simulation models of the behavior of the pollutants in the natural environment and in treatment systems, with a view to elaborating better ways of operating such systems and of forecasting the consequences of alternative schemes for managing and protecting the natural environment. In addition, the instrument will be used in a graduate-level course to teach students how to use it. Application accepted by Commissioner of Customs: October 1, 1996.

Docket Number: 96-105. Applicant: Arizona Science Center, 147 E. Adams Street, Phoenix, AZ 85004-2394. Instrument: Interactive Imaging System, Model Magicam. Manufacturer: Optech International Ltd., New Zealand. Intended Use: The instrument will be used as an educational tool in geology and biology exhibit halls to allow the visitor to use the system to further explore provided examples in each of the galleries. Application accepted by Commissioner of Customs: October 2, 1996.

Docket Number: 96-106. Applicant: The Johns Hopkins University, Department of Chemistry, 3400 Charles Street, Baltimore, MD 21218. Instrument: EPR Spectrometer, Model EMX 10/2.7. Manufacturer: Bruker Instruments, Inc., Germany. Intended Use: The instrument will be used for electron spin resonance measurements at room and variable temperatures

during investigations that include characterization of paramagnetic centers in biomolecules, organic compounds, inorganic coordination compounds and solid state materials, identification of photo- and redox-active sites and elucidation of reaction mechanisms. In addition, the instrument will be used for educational purposes in chemistry laboratory courses. Application accepted by Commissioner of Customs: October 2, 1996.

Docket Number: 96-108. Applicant: Centers for Disease Control & Prevention, Mailstop G-36, 1600 Clifton Road, N. E., Atlanta, GA 30333. Instrument: Mass Spectrometer, Model Reflex II. Manufacturer: Bruker Analytical, Germany. Intended Use: The instrument will be used to assess the molecular weight of the intact biopolymers and of synthetic intermediates employed in the syntheses and fragments generated from the biopolymers. Together, this information provides important evidence for the correct structure of the synthetic biotechnology products.

Application accepted by Commissioner of Customs: October 7, 1996.

Frank W. Creel,

Director, Statutory Import Programs Staff.

[FR Doc. 96-27771 Filed 10-29-96; 8:45 am]

BILLING CODE 3510-DS-P

The University of Texas, et al. Notice of Consolidated Decision on Applications for Duty-Free Entry of Scientific Instruments

This is a decision consolidated pursuant to Section 6(c) of the Educational, Scientific, and Cultural Materials Importation Act of 1966 (Pub. L. 89-651, 80 Stat. 897; 15 CFR part 301). Related records can be viewed between 8:30 A.M. and 5:00 P.M. in Room 4211, U.S. Department of Commerce, 14th and Constitution Avenue, N.W., Washington, D.C.

Comments: None received. Decision: Approved. No instrument of equivalent scientific value to the foreign instruments described below, for such purposes as each is intended to be used, is being manufactured in the United States.

Docket Number: 96-083. Applicant: The University of Texas at Austin, Austin, TX 78712. Instrument: Gas Composition Analyzer, Model Epison III. Manufacturer: Thomas Swan & Co., Ltd., United Kingdom. Intended Use: See notice at 61 FR 46782, September 5, 1996. Reasons: The foreign instrument provides non-invasive control of gas mixture ratios in a chemical vapor